

# EC Declaration of Product Conformity

The manufacturer established in the community;

**BM Polyco Ltd  
Crown Road  
Enfield  
EN1 1TX  
United Kingdom**

Certified to ISO 9001 & ISO 13485

declares that the new medical device described hereafter



## WHITE NITRILE GL901

Non-sterile, powder-free, nitrile disposable glove

is in conformity with the provisions of Council Directive 93/42/EEC and with National Standards transposing harmonised standards EN455-1, EN455-2, EN455-3, EN455-4 and is self-certified as a Class 1 non-sterile medical device

Freedom from holes: AQL 1.5  
Physical properties: 6N minimum

---

is in conformity with the provisions of Council Directive 89/686/EEC and where such is the case, with National Standards transposing harmonised standards EN374-1:2003, EN374-2:2003, EN374-3:2003 and EN420:2003.

is identical to the PPE which is the subject of EC certificate of conformity no. 1572 issued by SATRA Technology Centre, Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, United Kingdom (Notified Body No. 0321).

conforms to the provisions of Council Directive 89/686/EEC Article 11A under the supervision of SATRA (Notified Body No. 0321).

**MICRO ORGANISM  
HAZARDS  
EN374-2**



**CHEMICAL  
HAZARDS  
EN374-3**



# CE 0321



Complex Design

## User Information

This product does not contain natural rubber latex.  
Glove Care: Store below 40 °C away from direct sunlight.  
You are advised to retain this packaging for reference.  
Contains accelerators which may cause allergic reactions.

## Range Available: sizes and codes

Size	X-Small	Small	Medium	Large	X-Large
Code	GL9014	GL9011	GL9012	GL9013	GL9015

## Food Contact



GL901 have been tested in accordance with EN1186. They are suitable for use with all food categories in situations of short term, repeat contact.

Done at Enfield, 10/10/14

Bernard Garvey  
Technology Director